



Date: **27/10/2023**

NIQ No: **AIIMS/MG/Non-DPR/Neurosurgery/Rhoton Dissector Set/01/2023-24.**

Sub: - “NIQ for Supply of Rhoton Dissector Set” at AIIMS Mangalagiri.

**Notice Inviting Quotations**

On behalf of The Director, AIIMS Mangalagiri Quotations are invited with price bid and other documents from eligible Manufacturers/ Firms/ Companies/ Authorized Agents/ Distributors/ Dealers/ Supplier Agencies for “Supply of Rhoton Dissector Set” at AIIMS Mangalagiri, as per specified terms and conditions.

**I. Schedule of Requirement:-**

S.No	Description of items	Required Quantity
	Rhoton Dissector Set	01 Set (Set of 21 Items)
1.	RHOTON MICRO ROUND DISSECTOR- IMM ANGLED-190mm length #01	01 No.
2.	RHOTON MICRO ROUND DISSECTOR-2MM ANGLE-190mm length #02	01 No.
3.	RHOTON MICRO ROUND DISSECTOR-3MM ANGLED-190mm length #03	01 No.
4.	RHOTON ELEVATOR ANGLED - 1.5 WIDTH - 14MM DEPTH-190mm length #04	01 No.
5.	RHOTON ELEVATOR CURVED - 1.8 WIDTH - 190mm length #05	01 No.
6.	RHOTON SPATULA DISSECTOR SMALL - 1.5mm WIDTH-190mm length #06	01 No.
7.	RHOTON SPATULA DISSECTOR MEDIUM-2MM WIDTH-190mm length #07	01 No.
8.	RHOTON SPATULA DISSECTOR LARGE-3MM WIDTH-190mm length #08	01 No.
9.	RHOTON RIGHT ANGLE HOOK SEMI SHARP-1.5mm DEPTH-190mm length #09	01 No.
10.	RHOTON RIGHT ANGLE HOOK BLUNT - 1.8mm DEPTH-190mm length #10	01 No.
11.	RHOTON NEEDLE Angled 45 deg semi-Sharp 3.5mm Depth- 190mm length # 11	01 No.
12.	RHOTON NEEDLE STRAIGHT SHARP-190mm length #12	01 No.
13.	RHOTON MICRO CURETTE STRAIGHT-2mm tip – 190 mm length #13	01 No.
14.	RHOTON MICRO CURETTE STRAIGHT - 2.5mm tip - 190 mm length #13	01 No.
15.	RHOTON BALL DISSECTOR Straight - 1.2mm ball diameter - 190mm length #15	01 No.
16.	RHOTON BALL DISSECTOR Angled 90deg - 1.2mm ball diameter 2.5mm tip- 190mm length #16	01 No.

S.No	Description of items	Required Quantity
17.	RHOTON BALL DISSECTOR Angled 90deg - 1.2mm ball diameter - 5mm tip - 190mm length #17	01 No.
18.	RHOTON BALL DISSECTOR Angled 40deg-1.2mm ball diameter - 5mm tip - 190mm length #18	01 No.
19.	RHOTON BALL DISSECTOR Angled 40deg - 1.2mm ball diameter - 8mm tip- 190mm length #19	01 No.
20.	RHOTON INSTRUMENTS CASE	01 No.
21.	Shunt tube passer with blunt obturator tube O.D 4mm 62 CM long	01 No.

## **II. Specifications of Rhoton Dissector Set:**

- All the Rhoton dissector instruments should be made of corrosion-free surgical Titanium or better material and the bidder should submit an original OEM certificate for the same; Shunt tube passer instrument should be made of corrosion-free surgical grade stainless steel SS 400 grade or better and the bidder should submit an original OEM certificate for the same.
- All the firms are required to submit their catalog along with tender/quotation clearly marking the quoted item and no, with a highlighter.
- It should be of high quality and precision.
- It should be lightweight, strong, and durable.
- It should be non-magnetic.
- All instruments should be autoclavable at 134 degrees Celsius.
- The TC inlay should be welded to the instrument and not pasted.
- All instruments should be supplied from a single manufacturer (80% of the instruments should be from the same manufacturer).  
Undertaking from the principal manufacturer should be submitted for not manufacturing balance instruments.  
Rest 20% items should also meet quality standards as asked in the tender  
Note: MAF is required only for the principal manufacturer.
- Manufacturers should meet the International Certification as per ISO 13485.
- Instruments are to be supplied in autoclavable instrument boxes with silicone padding for Delicate instruments.
- Tolerance of +/- 10% is acceptable in dimension for all instruments.
- A warranty period of not less than 1 year for all items. The defective items should be repaired and replaced as required in the warranty period.
- All the instruments should be European CE with 4 digits notified body no or US FDA or BIS approved.

## **III. Terms and Conditions**

- The sealed quotation must have valid GST number
- Rates quoted must include all types of taxes and other possible expenses. No other charges will be considered.
- The payments will be made against delivery item with valid invoice and E-way bill through electronically viz RTGS/NEFT. The following information also to be mentioned in the quotation.
  - Name of the beneficiary:
  - Account No. of the beneficiary:
  - IFSC of the bank/branch:

**d). Delivery period:** Delivery period of the items shall be **30 days** from the issue of supply order. In the event of any deviation, the supply order may be cancelled.

e) The quotations will be **opened on 07/11/2023 03:00 PM** at Admin & Library Building, 2<sup>nd</sup> Floor, Room no.224 AIIMS Mangalagiri.

f). Vendors requested to attend at time of opening of bid.

**g) Penalty:** Penalty @ **0.50% per week** will levied for late delivery on supply order value of the item and will be deducted from the payment Bill. The maximum penalty will be **10 %** of total value against any of Purchase/Supply Order total value.

**h) Warranty Period:** Should be clearly mentioned in the quotation as per the OEM or company norms.

**i) Price Quotation:** price cannot exceed Rs.2.5 Lakhs.

**j) Price Validity:** The quoted prices should remain valid for a period of **180 days** from date of opening of NIQ.

**k).**The outer envelope should be super scripted as “**Supply of Rhoton Dissector Set**” at AIIMS Mangalagiri.

**IV.** The duly sealed and super scribed Quotations should reach **Procurement Cell, Admin & Library Building, 2<sup>nd</sup> Floor, Room no. 236, AIIMS Mangalagiri, Guntur-522503 by 15:00 PM on 06/11/2023.**

**Sd/-**  
**F I/c (Procurement)**  
**For and on behalf of Director**  
**AIIMS, Mangalagiri**