



भारत सरकार/Government of India  
स्वास्थ्य और परिवार कल्याण मंत्रालय/ Ministry of Health and Family Welfare  
प्रधानमंत्री स्वास्थ्य सुरक्षा योजना/PMSSY  
अखिल भारतीय आयुर्विज्ञान संस्थान/All India Institute of Medical Sciences  
मंगलगिरि, आंध्र प्रदेश/Mangalagiri, Andhra Pradesh

[www.aiimsmangalagiri.edu.in](http://www.aiimsmangalagiri.edu.in)

No.: AIIMS MG/Procurement/ PAC /2021-22/ Continuous Non Invasive Hemoglobin Monitor


Date: 11/05/2022

**CALL FOR OBJECTION**

AIIMS Mangalagiri has a requirement of **Continuous Non Invasive Hemoglobin Monitor** for the department of Anaesthesia. In this regard comments/ objection, if any are invited before declaring the article as proprietary in nature.

In this regard, the proposal submitted by **M/s New Path Solutions.**, Shed-2 Mawai village, Faridabad, Haryana (agent of **M/s. Masimo Medical Technologies India Pvt Ltd, Bangalore, Karnataka** for and behalf of **M/s. Masimo Corporation U.S.A** of this product) along with Proprietary Article Certificate are attached and duly uploaded on Institute website for open information to submit objections/comments, if any from any manufacturer/ supplier before declaring the article of the said equipment/items to be procured as proprietary.

The objections/comments should be sent to the O/o Faculty I/C Procurement, 4<sup>th</sup> floor, Nursing College building, AIIMS Mangalagiri, Guntur, Andhra Pradesh 522503 in a sealed envelope with the above reference on or before 21/05/2022, 3:00 PM, i.e., 10 days from today failing which it will be presumed that no other manufacture/vendor is having any objections/comments to offer and the case will be decided on merits.

  
Asst. FI/C Procurement  
For Director, AIIMS Mangalagiri

**Enclosure:**

1. PAC of Department
2. PAC of Manufacturer
3. Manufacturer Authorization form

Email: [Procurement@aiimsmangalagiri.edu.in](mailto:Procurement@aiimsmangalagiri.edu.in)



P-3 FORM

(to be attached with P-2 form for Proprietary items)

AIIMS Mangalagiri

PROPRIETARY ARTICLE CERTIFICATE

It is certified that the items Continuous Non Invasive Hemoglobin Monitor Make: Masimo Model: Radical - 7 required in the P-2 form should be purchased from Masimo. To the best of my knowledge Masimo are the sole manufacturer/ agents of the sole manufacturers

Similar items manufactured by other firm(s) shall not be suitable for our purpose for the following reasons:

1. Rainbow Signal Extraction Technology that allows continuous non-invasive measurement of Hemoglobin (SpHb) is proprietary technology to Masimo.

**Dr. VANSIDHAR CHAMALA**  
KMC Regd. No : 61248  
Asst. Professor  
Dept. of Anesthesiology  
AIIMS, MANGALAGIRI

Date: 30-06-2021

Designation: Assistant Professor

Department: Anesthesiology

Recommend at on:

Signature of Head of Department of Anesthesiology  
HOD, Department of Anesthesiology  
All India Institute of Medical Sciences  
Mangalagiri, Andhra Pradesh.

N.B.: The indenter before recording the above certificate should satisfy himself that the article is genuinely of proprietary nature manufactured under patent laws.





Masimo Medical Technologies India Pvt. Ltd  
10/2 Millers Road, Millers Boulevard, 2nd Floor  
Bangalore 560022, Karnataka State

May 31, 2018

**TO WHOM IT MAY CONCERN**

This is to certify that the following technologies, and parameters are unique to pulse-oximeters and Co-oximeters manufactured by **Masimo Corporation**, are protected by patents, trademarks and or trade secrets as appropriate, and are proprietary to Masimo Corporation:

1. **Signal Extraction Technology™ (SET™)**: Proprietary technology protected by patents, trademark, and trade secrets, that allow the measurement of oxygen saturation in the blood, in spite of motion (Otherwise know as 'read through motion') and low perfusion up to 0.02 Perfusion Index (PI)

2. **Rainbow SET™ or Rainbow Technology**: Proprietary technology protected by patents, trademark, and trade secrets, that allow the non-invasive and continuous measurement of the following parameters:

- i. Hemoglobin (SpHb)
- ii. Carboxyhaemoglobin (SpCO)
- iii. Methaemoglobin (SpMet)
- iv. Pleth Variability Index (PVI)
- v. Acoustic Respiration Rate (RRa)

**The combination of these measurements in one device is unique to Masimo, and is available as required in Masimo's Radical & Pulse Oximeter, Rad9, Rad97, and Rad 97 with NIBP**

Further, to the best of our knowledge, no other company or manufacturer in the market, other than those who have licensed the technology from Masimo, produces, or can offer **SET™, Rainbow SET™, Read through Motion, or Read through Low Perfusion up to 0.02 Perfusion Index**, technologies and capabilities.

For and on Behalf of Masimo Corporation:



**Bharat D. Monteiro**  
Country Manager - India & The Sub-continent





Masimo Medical Technologies India Pvt. Ltd.  
Masimo Medical Technologies India Pvt. Ltd.  
Masimo Medical Technologies India Pvt. Ltd.

Date: 10-12-2021

SECTION – XII MANUFACTURER'S  
AUTHORISATION FORM

The Director  
All India Institute of Medical Sciences  
Manglagiri, India.

Dear Sir,

Ref: Your Bid document No GEM/2021/B/1692744 for Continuous non-invasive hemoglobin Monitor.

We, Masimo who are proven and reputable manufacturers Of Continuous non-invasive hemoglobin Monitor Having factories at USA, hereby authorize New Path Solutions to submit a bid, process the same further and enter into a contract with you against your requirements contained in the above referred Bid documents for the above goods manufactured by us.

We also state that we are not participating directly in this bid for the following reason(s): Company has decided to quote the Tender through our Authorized dealer.

We further confirm that no supplier or firm or individual other than Messrs. New Path Solutions having address Shed-2, Mawai Village, Faridabad Haryana is authorized to submit a bid, process the same further and enter into a contract with you against your requirement as contained in the above referred Bid documents for the above goods manufactured by us. We also hereby extend our full warranty, CAMC as applicable as per clause 15 of the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this Bid document.

We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorized agent and the spares for the equipment shall be available for at least 10 years from the date of supply of equipment.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly.

Yours faithfully,  
For Masimo Medical Technologies India Pvt Ltd.,

